



OIML CERTIFICATE OF CONFORMITY

To Xiamen Pinnacle Electrical Co., Ltd.

This certificate attests the conformity of the type of the

Barcode Label Scale LS2615E, LS2615EC, LS21530E,

LS21530EC, LS215E, LS215EC with the requirements

of the Recommendation R76(1992) of the

International Organization of Legal Metrology(OIML)

The issuing authority

Office of OIML Affairs

中国秘书处

July 25, 2007

The OIML member

蒲长城

July 25, 2007

Member State
P.R.China

OIML Certificate No.
R76/1992- CNI-07.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: Xiamen Pinnacle Electrical Co., Ltd.
Address: 4F, Guangxia Building, Torch High-Tech Zone, Xiamen, Fujian Province, 361006 P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Barcode Label Scale

Type	LS2615E LS2615EC	LS21530E LS21530EC	LS215E LS215EC
Accuracy class	III		
Max (Max ₁ /Max ₂)	6kg/15kg	15kg/30kg	15kg
Min	40g	100g	100g
e (e ₁ /e ₂)	2g/5g	5g/10g	5g
n (n ₁ /n ₂)	3000/3000	3000/3000	3000
Tare balancing range(subtractive)	≤ 9.995kg	≤ 9.990kg	≤ 7.990kg
Temperature range	0°C ~ 40°C		
Humidity range	10%RH ~ 85%RH		

The numbers of the associated test reports and the included pages:
NIM-LSmc-0601 that includes 61 pages.